

TCLP EPA 1311 / SPLP EPA 1312			
Method	Analyte	Reporting Limit mg/L	Regulatory Level mg/L
SW6020	Arsenic	0.050	5.0
SW6020	Barium	1.000	100.0
SW6020	Cadmium	0.050	1.0
SW6020	Chromium	0.050	5.0
SW6020	Lead	0.050	5.0
SW6020	Mercury	0.002	0.2
SW6020	Selenium	0.050	1.0
SW6020	Silver	0.050	5.0
SW8081	Chlordane	0.0067	0.030
SW8081	Endrin	0.0003	0.020
SW8081	Gamma-BHC (Lindane)	0.0003	0.400
SW8081	Heptachlor	0.0003	0.008
SW8081	Heptachlor Epoxide	0.0003	0.008
SW8081	Methoxychlor	0.0003	10.000
SW8081	Toxaphene	0.0067	0.500
SW8151	2,4-D	0.01	10.0
SW8151	2,4,5-TP (Silvex)	0.001	1.0
SW8260	1,1-Dichloroethene	0.1	0.70
SW8260	1,2-Dichloroethane	0.1	0.50
SW8260	1,4-Dichlorobenzene	0.1	7.50
SW8260	2-Butanone (MEK)	1.0	200.00
SW8260	Benzene	0.1	0.50
SW8260	Carbon tetrachloride	0.1	0.50
SW8260	Chlorobenzene	0.1	100.00
SW8260	Chloroform	0.1	6.00
SW8260	Tetrachloroethene	0.1	0.70
SW8260	Trichloroethene	0.1	0.50
SW8260	Vinyl chloride	0.1	0.20
SW8270	2,4,5-Trichlorophenol	0.01	400.00
SW8270	2,4,6-Trichlorophenol	0.01	2.00
SW8270	2,4-Dinitrotoluene	0.01	0.13
SW8270	2-Methylphenol	0.01	200.00
SW8270	3&4-Methylphenol	0.01	200.00
SW8270	Hexachlorobenzene	0.01	0.13
SW8270	Hexachlorobutadiene	0.01	0.50
SW8270	Hexachloroethane	0.01	3.00
SW8270	Nitrobenzene	0.01	2.00
SW8270	Pentachlorophenol	0.02	100.00
SW8270	Pyridine	0.01	5.00