

Metals Elements by EPA6020 & 200.8 in Water Matrix					
Analyte	PP	RCRA	TAL	RL ug/L	LOD ug/L
Aluminum			X	100.00	50.00
Antimony	X		X	1.00	0.50
Arsenic	X	X	X	1.00	0.50
Barium		X	X	1.00	0.50
Beryllium	X		X	1.00	0.50
Boron				20.00	10.00
Cadmium	X	X	X	1.00	0.50
Calcium			X	100.00	50.00
Chromium	X	X	X	1.00	0.50
Cobalt			X	1.00	0.50
Copper	X		X	1.00	0.50
Iron			X	100.00	50.00
Lead	X	X	X	1.00	0.50
Lithium				5.00	2.50
Magnesium			X	100.00	50.00
Manganese			X	1.00	0.50
Mercury	X	X	X	0.20	0.10
Molybdenum				1.00	0.50
Nickel	X		X	1.00	0.50
Potassium			X	100.00	50.00
Selenium	X	X	X	1.00	0.50
Silver	X	X	X	1.00	0.50
Sodium			X	100.00	50.00
Thallium	X		X	1.00	0.50
Tin				1.00	0.50
Vanadium			X	5.00	2.50
Zinc	X		X	20.00	10.00
Zirconium				10.00	5.00